

**1. Part Number:** 8015H-170S

**2. Photograph:**



### 3. Product Information:

<b>Description:</b>	SteriWare® Micro Spatula HDPE - Sterile
<b>Use:</b>	Single use, disposable spatula
<b>Material of Construction:</b>	High Density Polyethylene (HDPE) <ul style="list-style-type: none"> <li>- FDA acceptable CFR 177.1520</li> <li>- EU Regulations 10/2011 compliant</li> <li>- EC Regulations 1935/2004 compliant</li> <li>- BSE/TSE free</li> </ul>
<b>Method of Construction:</b>	Single piece, injection moulded
<b>Nominal Volume:</b>	2ml
<b>Overall Length:</b>	175mm
<b>Spatula Blade Length:</b>	40mm
<b>Blade Width:</b>	13mm
<b>Nominal Weight:</b>	4g
<b>Moulding Environment:</b>	Class 100,000 Cleanroom
<b>Primary Packing Environment:</b>	Class 10,000 Cleanroom
<b>Individually Bagged?</b>	Yes (heat sealed PE/PET bag)
<b>Method of Sterilisation:</b>	Gamma Irradiation (25-45kGy)
<b>Number of Scoops per Box:</b>	100
<b>Recommended Storage Conditions:</b>	Dry and ambient temperature
<b>Shelf Life:</b>	5 years from the date of manufacture
<b>ATEX:</b>	This product has not been ATEX tested

#### 4. Change Control:

Sampling Systems operates a policy of continual improvement.

To be kept informed of all major changes to any product in the SteriWare range please email your contact details to [change.control@sampling.com](mailto:change.control@sampling.com)

#### 5. Country of Origin:

Made and packed in the UK.

#### 6. Repeated Use:

This is a single-use, disposable device. Repeated use will lead to failure.

#### 7. Approval:

Signed:



Name: Kuldeep Flora

Position: Quality Manager

Date: 9-Jul-24